

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	S-2302, 25 mg Substrate	EXPIRATION DATE:	2023-03
PRODUCT NUMBER:	82034039		
LOT NUMBER:	N1190264		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

PREPARED BY: Maria Oettinger
PRINT

Maria Oettinger
SIGNATURE

DATE: 1-23-2020

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz
PRINT

Debra Lee Harz
SIGNATURE

DATE: 1/24/2020

TITLE: QC Technical Supervisor