

TECO medical

TECO® Intact Proinsulin ELISA KIT

For research use only

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

REF
LOT

TE1011
026A



2021-09-30

Refer to Product Insert for further Information
Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts
Se référer au Protocole du Dosage pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

Calibrators could vary
from batch to batch

Standards können von Batch
zu Batch unterschiedlich sein

La concentration des
standards peut varier selon les lots

Calibrators	Δ pmol/L WHO 09/296	\bar{x} Extinction 450 nm	\bar{x} Extinction 405 nm
A	0.0	0.000	0.000
B	3.8	0.085	0.027
C	9.0	0.271	0.088
D	16.2	0.528	0.168
E	48.7	1.384	0.443
F	113.5	-	1.005
L - Control 1	11.2 (7.3 - 15.8)	0.350	0.112
H - Control 2	27.2 (20.4 - 33.9)	0.886	0.283

ITEM

LOT

EXP

① Specific Intact Proinsulin Antibody coated wells	EBX-DZ-042	2021-09
② Blocking buffer	EBX-LA-034	2021-09
③ Antibody HRP Conjugate	EBX-AZ-2015	2021-09
④ TMB Substrat	GZZ-IE-032	2022-02
⑤ Concentrated wash solution	EBX-HZ-030	2021-09
⑥ Stop Solution	GZZ-KZ-022	2023-03
A Calibrator A	EBX-BZ-023	2022-03
B Calibrator B	EBX-BZ-023	2022-03
C Calibrator C	EBX-BZ-023	2022-03
D Calibrator D	EBX-BZ-023	2022-03
E Calibrator E	EBX-BZ-023	2022-03
F Calibrator F	EBX-BZ-023	2022-03
L Control 1	EBX-BZ-023	2022-03
H Control 2	EBX-BZ-023	2022-03

TECOmedical

Qc release

Date

2020-06-02