Certificate of Analysis

READS
VWF Antigen ELISA

Kit Lot: VF-220
Plate Lot #: VW-069
Product #: 034-001
Expiration Date: 2021-11-19

Quality Control Results

VWF Antigen
Testing Date: 2020-07-02
Manufacture Date: 2020-05-19

Attribute	Value	Specification	Result by Assay
Reference Plasma 100% OD	110 % of Normal	≥ 1.0	1.184
Reagent Background O.D.		≤ 0.100	0.026
Reference Curve R-Squared		≥ 0.95	0.99758

Control Recovery	Label Range	Result by Assay
Normal Control Control Name N/A	102 - 153 %	142.0 %
Deficient Control Control Name N/A	19 - 43 %	34.6 %

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: Souda Truong

Signed: _____ **2020-07-09**
Title: Senior Quality Assurance Associate

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519