

	<p style="text-align: center;"><b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b></p>	<p style="text-align: center;">DQ6V08.03</p>
-----------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------	----------------------------------------------

**Dapttin TC 6x10 mL**

REF      5035090  
 LOT      6V91BB0.01  
      2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		6.6 – 7.2	6.9	OK	
Verwendbarkeit Usability Ceveron® alpha	Daptlin TC, sec	CCN LOT 1P71B01	31.0 – 38.2 sec	33.8	OK
		CCA LOT 3P71B01	73.5 – 90.7 sec	74.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von  
Prepared by

Riham EL KAMMAR  
*El Kammar Riham*

11. SEP. 2019

Datum / Date:

Freigabe von  
Approved by

Eva Maria HAGE  
*Eva Maria Hage*

11. SEP. 2019

Datum / Date: