



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3C00.02

TECHNOVIEW UFH Control L 5x1 mL

REF 5090072

LOT 3C01C00.01

 2022-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® Alpha	Anti-Xa UFH	0.16 – 0.30 IU/mL	0.25	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Suzana GRESAK PODAKAR

12. OKT. 2020

Datum / Date:

Freigabe von
Approved by


Stefan STEINER

13. OKT. 2020

Datum / Date: