

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
-----------------------------------------------------------------------------------	-------------------------------------------------------------	-----------

TECHNOVIEW® Apixaban CONT Low 5x1 ml

REF 5090271
LOT 3901C00.02
 2022-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		76.6 – 142.2 ng/mL	109.4	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	76.6 – 142.2 ng/mL	104.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

14. OKT. 2020
Datum / Date:

Freigabe von
Approved by


Christoph ENNE

16. OKT. 2020
Datum / Date: