



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0Z00.02

TECHNOCHROM[®] FXIII 3 x 3 mL

REF 5360010
LOT 0Z05C00.01
⌚ 2021-09-30

Inhalt	Reagenz	LOT	⌚
3 x 3 mL	FXIII Activator Reagent	2004100/1	2021-09-30
3 x 3 mL	FXIII Detection Reagent	2004100/2	2021-09-30
3 x 3 mL	FXIII NADPH	2004100/3	2021-09-30
3 x 1 mL	FXIII Inhibitor Reagent	2004100/4	2021-09-30
1 x 6 mL	Stabilizer solution	2004100/5	2021-09-30

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P02C00	60.8 – 101.3 %	81.4
	Coagulation Control A Lot: 3P91B00	24.0 – 40.0 %	32.9

Durchführung von
Prepared by

Michael KAFKA

26. NOV. 2020

Datum / Date:

Freigabe von
Approved by

HOLOVICZ

26. NOV. 2020

Datum / Date: