

	ANALYSENZERTIFIKAT	DQRK00.01
	CERTIFICATE OF ANALYSIS	Att.05
		RK02C00.01

TECHNOZYM vWF:CBA Collagen Typ I ELISA Kit 96T.

REF	5450311
LOT	RK02C00.01
	2022-02-28

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 2	2ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 3	3ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 4	4ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 5	5ND01C0	2022-02-28
1x	0.5 mL	vWF High Control	AND01C0	2022-02-28
1x	0.5 mL	vWF Low Control	BND01C0	2022-02-28
1x	12 mL	TMB Substrate	RKN02C0	2022-06-30
1x	12 mL	Stop solution	RKK02C0	2022-06-30
1x	80 mL	Wash buffer concentrate	RNA02C0	2022-07-31
1x	90 mL	Incubation buffer	RNB02C0	2022-06-30
1x	-	vWF:CBA Plate Collagen Typ I	RNM01C0	2022-04-30
1x	0.3 mL	CBA Conjugate	RNH01C0	2022-09-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Ca11-Ca15	≥ 0.7	2.2	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND01C0	0.76 - 1.26 IU/mL	0.99 IU/mL	OK
	vWF Low Control LOT BND01C0	0.17 - 0.39 IU/mL	0.28 IU/mL	OK