
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

## TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF      5450451  
LOT      RP0CC0S.01  
      2021-11-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 6	Adamts13 INH Plate	RPS03CS	2022-03-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV92B0	2021-11-30
1x 80mL	Wash buffer Concentrate	RNA02C0	2022-07-31
1 x 90mL	Incubation buffer	RNB02C0	2022-06-30
1 x 12mL	TMB Substrate	RKN01C0	2022-01-31
1 x 12mL	Stop solution	RKK02C0	2022-06-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.60	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 - 93.94 U/mL	76.08	OK
	Low Control (RPZ92B0)	15.11 - 35.27 U/mL	28.18	OK

**Durchführung von**  
**Prepared by**

Michael KAFKA



02. NOV. 2020

Datum / Date:

**Freigabe von**  
**Approved by**

Fikret NASUFI



03. NOV. 2020

Datum / Date: