

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------


TECHNOZYM® ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT01C0.02
 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 – 93.94 U/mL	77.09	OK
	Low Control (RPZ92B0)	15.11 – 35.27 U/mL	27.25	OK

**Durchführung von
Prepared by**

Michael KAFKA 

02. NOV. 2020
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

03. NOV. 2020
Datum / Date: