



TECHNOFLUOR FXIII Activity

REF	5800200
LOT	FB01C00.01
	2022-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
2x	2mL	TECHNOFLUOR FXIII Reagent 2mL	FBA01C0	2022-04-30
2x	2mL	TECHNOFLUOR FXIII Substrate 2mL	FBB01C0	2022-04-30
2x	3mL	TECHNOFLUOR FXIII Buffer 3mL	FBC01C0	2022-04-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP
	Vial label	according to packaging SOP
	Vial barcode	according to packaging SOP
	Box label	according to packaging SOP
	Box barcode	according to packaging SOP
	Package insert	present
	Batch table	present
Kontrollwiederfindung Ceveron s100	Control N LOT 1P02C00	60,0 - 100,0 % 87,7 % OK
Recovery of controls	Control A LOT 3P91B00	25,5 - 42,5 % 33,3 % OK

Durchgeführt von
Prepared by


Martina LEITNER

27-04-2022
Datum / Date:

Freigabe von
Approved by


Margaret GRIFFITHS

27-04-2022
Datum / Date: