

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME: <u>S-2222, 25 mg Substrate</u>	EXPIRATION DATE: <u>2023-07</u>
PRODUCT NUMBER: <u>82031639</u>	
LOT: <u>N0303080</u>	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	Pass
FREE p-NITROANILINE	0	0.5	mol %	0	Pass

PREPARED BY:

Maria Oettinger
PRINT

Maria Oettinger
SIGNATURE

DATE: 4-20-20

TITLE: QC Analyst III

APPROVAL SIGNATURE:

Debra Harz
PRINT

Debra Harz
SIGNATURE

DATE: 4-20-20

TITLE: QC Technical Supervisor

QCM00820316-2 Rev. 00