

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2023-06
PRODUCT NUMBER:	82141339		
LOT:	N0605413		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.6	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.4	PASS

PREPARED BY: Bhumi Vasoya Bhumi Vasoya DATE: 07/23/2020
 PRINT SIGNATURE

TITLE: Lc Analyst

APPROVAL SIGNATURE: Debra Lee Harz Debra Lee Harz DATE: 07/23/2020
 PRINT SIGNATURE

TITLE: QC Tech Supervisor