



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRXXAZ

TECHNOZYM® u-PA ELISA Kit 96T

REF TC12010

LOT 120101C0.01

🕒 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x0.5mL	AQ-Standard u-PA conc.	RAQ01C0	2022-10-31
1x12mL	KK stop solution for TMB	RKK02C0	2022-06-30
1x12mL	TKN MB Substrat	RKN01C0	2022-01-31
1x80mL	NA Wash buffer + thimerosal	RNA01C0	2022-02-28
1x	GU-anti u-PA coated plate	RGU71B1	2021-05-31
1x	Biotinylated Detecting Antibody	RAH01C0	2021-05-31
1x	Streptavidin-HRP	RAK01C0	2021-05-31
2x12mL	DD-POX Dilution Buffer PBS	RDD01C0	2022-02-28
1x20mL	DC-Dil.Buffer (u-PA) w EDTA	RDC01C0	2022-10-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 0.7	1.21 OK
	Calibrator 6	< 0.2	0.01 OK

Durchführung von
Prepared by

Michael KAFKA

29. OKT. 2020
Datum / Date:

Freigabe von
Approved by

Christoph ENNE

29. OKT. 2020
Datum / Date: