



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRD00AZ

TECHNOZYM[®] PAI-1 ELISA Kit 96T.

REF TC12075

LOT RD01C00.02

2021-10-31

Inhalt Content	Reagenz Reagent	LOT	
1x12mL	KK stop solution for TMB	RKK01C0	2022-01-31
1x0.3mL	KL PAI POX For TMB	RKL01C0	2022-03-31
1x12mL	KN TMB substrate	RKN01C0	2022-01-31
1x80mL	NA wash buffer + thimerosal	RNA01C0	2022-02-28
1x90mL	NB Incubation buffer	RNB93B0	2021-10-31
1x0.5mL	BQ Calibrator 1 PAI-1	1BQ91B0	2021-11-30
1x0.5mL	BQ Calibrator 2 PAI-1	2BQ92B0	2021-11-30
1x0.5mL	BQ Calibrator 3 PAI-1	3BQ91B0	2021-11-30
1x0.5mL	BQ Calibrator 4 PAI-1	4BQ91B0	2021-11-30
1x0.5mL	BQ Calibrator 5 PAI-1	5BQ92B0	2021-12-31
1x	GQ-anti PAI-1 coated plate	RGQ01C0	2022-03-31
1x0.5mL	ABQ-PAI-1 High Control	ABQ92B0	2021-11-30
1x0.5mL	BBQ-PAI-1 Low Control	BBQ91B0	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	1.0 – 2.5	1.38	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (ABQ92B0)	25.2 – 42.0 ng/mL	26.21	OK
	Low Control (BBQ91B0)	2.8 – 6.4 ng/mL	3.84	OK

Durchführung von
Prepared by
Michael KAFKA

12. AUG. 2020

Datum / Date:

Freigabe von
Approved by
Christoph ENNE

13. AUG 2020

Datum / Date: