



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0800.01

TECHNOCHROM® anti-Xa Kit 80T.

REF 5340250

LOT 0802C00.01

⌚ 2022-03-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x20 mL	R1 anti Xa Buffer	9R91B00	2024-01-31
1x4 mL	R2 FXa	7K91B01	2022-03-31
1x4 mL	R3 Substrat	8A01C00	2022-07-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Anti-Xa Apixaban Ceveron® alpha	Apixaban Cont L LOT 390100	76.6 – 142.2 ng/mL	113.1 OK
	Apixaban Cont H LOT 3D01C00	214.7 – 357.9 ng/mL	299.7 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von  
Prepared by

Stefan STEINER

26. AUG. 2020

Datum / Date:

Freigabe von  
Approved by

Suzana GRESAK PODAKAR

26. AUG. 2020

Datum / Date: