



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP0BC00.02
⌚ 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 INH Plate	RPS07C0	2022-09-30
1 x 0.3mL	Adamts13 INH Conjugate	RPV92B0	2021-11-30
1x 80mL	Wash buffer Concentrate	RNA02C0	2022-07-31
1 x 90mL	Incubation buffer	RNB02C0	2022-06-30
1 x 12mL	TMB Substrate	RKN02C0	2022-06-30
1 x 12mL	Stop solution	RKK02C0	2022-06-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.55	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 - 93.94 U/mL	77.17	OK
	Low Control (RPZ92B0)	15.11 - 35.27 U/mL	26.73	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

03. NOV. 2020

Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

04-Nov-2020

Datum / Date: