

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6PT00AZ
---	---	-----------


**TECHNOZYM® ADAMTS-13 INH Calibrator Set 5x0.5mL**

REF      5450461  
LOT      6PT02C0.01  
      2022-09-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT01C0	2022-09-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT01C0	2022-09-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT01C0	2022-09-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT01C0	2022-09-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT01C0	2022-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.39	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY01C0)	54.04 – 90.06 U/mL	79.03	OK
	Low Control (RPZ01C0)	14.50 – 33.82 U/mL	26.60	OK


**Durchführung von  
Prepared by**

Michael KAFKA   


---

09. DEZ. 2020  
Datum / Date:

**Freigabe von  
Approved by**

Fikret NASUFI   


---

09. DEZ. 2020  
Datum / Date: