

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM[®] ADAMTS-13 INH Control Set 2x0.5mL

REF	5450463
LOT	7PT02C0.01
	2022-09-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY01C0	2022-09-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ01C0	2022-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (RPY01C0)	54.04 – 90.06 U/mL	79.03	OK
	Low Control (RPZ01C0)	14.50 – 33.82 U/mL	26.60	OK

**Durchführung von
Prepared by**

Michael KAFKA 

09. DEZ. 2020

Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

09. DEZ. 2020

Datum / Date: