



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

TECHNOTHROMBIN<sup>®</sup> TGA Kit 3x16 T.

REF 5006010

LOT 0T11C00.01

2021-11-30

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA CAL BUF	9P02C00	2022-08-31
3x1.5mL	TGA SUB	8F01C00	2022-02-28
1x0.5mL	TGA RC Low	7D01C00	2022-02-28
1x0.5mL	TGA RC High	7F01C00	2022-11-30
1x1.5mL	TGA RD	7G01C00	2022-11-30
1x0.5mL	TGA CAL	5T02C	2022-03-31
1x1mL	TGA Control High	1Z03C00	2021-11-30
1x1mL	TGA Control Low	2805C00	2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1Z03C00	153 – 271	178.3	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2805C00	102 – 284	173.9	OK
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1Z03C00	461 – 768	593.2	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2805C00	247 – 412	392.2	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1Z03C00	407 – 679	497.5	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2805C00	488 - 814	651.6	OK



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

Durchführung von  
Prepared by

Michael KAFKA

A handwritten signature in blue ink, appearing to be "MK", written over a horizontal line.

15. JAN. 2021

Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

A handwritten signature in blue ink, appearing to be "SSt", written over a horizontal line.

15. JAN. 2021

Datum / Date: