



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ7F00.02

Att.03

7F01C00.01

TECHNOTHROMBIN TGA RC High 5 x 0,5 mL

REF 5006214
LOT 7F01C00.01
⌚ 2022-11-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Cap color	turquoise	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	white colored plug of lyophilized material	confirmed	OK
	Appearance after reconstitution	clear liquid	confirmed	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1203C00	461.0 - 768.0 nM	528.2 nM	OK
	TGA Control Low (nM Peak Thrombin) LOT 2805C00	247.0 - 412.0 nM	348.6 nM	OK

Durchgeführt von
Prepared by

Michael KAFKA

18. JAN. 2021

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

18. JAN. 2021

Datum / Date: