

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7G00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RD Reagenz 5x1.5mL

REF 5006220
LOT 7G01C00.01
 2022-11-30

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA RD Reagenz	7G01C00	2022-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z03C00	407 - 679	453	OK
	TGA Control Low (nM Peak Thrombin) LOT 2805C00	488 – 814	591	OK

Durchführung von
Prepared by

Michael KAFKA 

18. JAN. 2021

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

18. JAN. 2021

Datum / Date: