

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ2800AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA Control Low 5x1mL

REF      5006330  
LOT      2805C00.01  
      2021-12-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1mL	TGA Control Low	2805C00	2021-12-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D01C00	102 – 284	190 OK
	TGA RC High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F01C00	247 – 412	349 OK

**Durchführung von  
Prepared by**

Michael KAFKA 

18. JAN. 2021

Datum / Date:

**Freigabe von  
Approved by**

Lucas HOLOVICZ 

18. JAN. 2021

Datum / Date: