
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0500.01
---	---	-----------


**Technochrom AT III manual method Kit ~39 T.**

REF      5340225  
LOT      0502C00.01  
      2021-05-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x	Substrate Th-1 10µmol QC	8B02C00	2023-03-31
1x	Technochrom ATIII Reagent A1 20µmol	7U61B02	2021-05-31
1x	Technochrom ATIII Puffer A1 100mL	9871B00	2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	CCN <span style="border: 1px solid black; padding: 2px;">LOT</span> 1P02C00	75.8 – 126.3%	90.8	OK
	CCA <span style="border: 1px solid black; padding: 2px;">LOT</span> 3P91B00	33.8 – 56.3%	43.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Laydrim NASUFI

25. SEP. 2020  
Datum / Date:

**Freigabe von**  
**Approved by**

  
Eva Maria HAGER

25. SEP. 2020  
Datum / Date: