



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQxND0.02

Att.20

6NG01C0.01

TECHNOZYM® vWF:Ag Calibrator Set 5x0.5 mL

REF	5450210
LOT	6NG01C0.01
	2021-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	⌚
1x	0.5mL	HFP: ND Calibrator 1 vWF	1ND91B0	2021-05-31
1x	0.5mL	HFP: ND Calibrator 2 vWF	2ND91B0	2021-05-31
1x	0.5mL	HFP: ND Calibrator 3 vWF	3ND91B0	2021-05-31
1x	0.5mL	HFP: ND Calibrator 4 vWF	4ND91B0	2021-05-31
1x	0.5mL	HFP: ND Calibrator 5 vWF	5ND92B0	2021-06-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch table	present confirmed OK
Verwendbarkeit Usability	Δ OD (Cal 1 - Cal 5)	\geq 0.7 1.4 OK
	vWF High Control LOT AND91B0	0.92 - 1.54 IU/mL 1.19 IU/mL OK
	vWF Low Control LOT BND91B0	0.22 - 0.50 IU/mL 0.39 IU/mL OK

Durchgeföhrt von
Prepared by

Michael KAFKA

07. FEB. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

07. FEB. 2020

Datum / Date: