

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	<u>Antithrombin 10 x 10 IU</u>	EXPIRATION DATE:	<u>2022-12</u>
PRODUCT NUMBER:	<u>82072039</u>		
LOT NUMBER:	<u>N0605411</u>		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
ANTITHROMBIN, 10 IU (654023)					PASS/FAIL
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV.

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 7-16-20
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 7-16-20
 PRINT SIGNATURE

TITLE: Qc Technical Supervisor