

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE: 2022-07
PRODUCT NUMBER:	82265063	
LOT	N0806743	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION	
	MIN	MAX				
Analysis with COATEST APC R:						
APT Time						
	<u>CaCl₂</u>	28	34	seconds	32	PASS
	<u>APC/CaCl₂</u>	90	--	seconds	134	PASS
APC Ratio		3.5	4.4	--	4.2	PASS

Analysis with COATEST APC R V:						
APT Time						
	<u>CaCl₂</u>	--	38	seconds	34	PASS
	<u>APC/CaCl₂</u>	75	--	seconds	98	PASS
APC Ratio		2.7	3.2	--	2.9	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Maria Oettinger
PRINT

Maria Oettinger
SIGNATURE

DATE: 9/15/2020

9-15-20

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.m. Patel

SIGNATURE

DATE: 9/15/2020

9-15-20

TITLE: Senior QC Analyst

QCM00C00423L-1, Rev. 01