

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2023-07
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0705719		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: BHUMI VASOYA
PRINT

Bhumi Vasoya
SIGNATURE

DATE: 08/06/2020

TITLE: QC ANALYST

APPROVAL SIGNATURE: D.Tencza
PRINT

D.Tencza
SIGNATURE

DATE: 8/6/2020

TITLE: Senior QC Analyst