



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

0201C00.01

## Technoview® Argatroban CAL Set 5x1mL

REF 5090140  
LOT 0201C00.01  
⌚ 2022-06-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	⌚
1x	1mL	Argatroban CAL 1	0201C00 5Q51B01	2022-06-30
1x	1mL	Argatroban CAL 2	5R51B01	2022-06-30
1x	1mL	Argatroban CAL 3	5S51B01	2022-06-30
1x	1mL	Argatroban CAL 4	5U51B01	2022-06-30
1x	1mL	Argatroban CAL 5	5V51B01	2022-06-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	Technoview® Argatroban CONT L LOT 3151B02	0.48 - 0.90 µg/mL	0.70 µg/mL	OK
	Technoview® Argatroban CONT H LOT 3251B02	0.90 - 1.50 µg/mL	1.20 µg/mL	OK

Durchgeführt von  
Prepared by

Stefan STEINER

07. JULI 2020

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

07. JULI 2020

Datum / Date: