



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

3251B02.01

Technoview® Argatroban CONT High 5x1mL

REF	5090144
LOT	3251B02.01
	2022-06-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Cap color	red	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
Appearance after reconstitution	buff colored liquid	confirmed	OK	
Virologie / Virology The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	TECHNOVIEW® Argatroban LOT 0X01C00	0.90 - 1.50 µg/mL	1.20 µg/mL	OK

Durchgeführt von
Prepared by

Stefan STEINER

07. JULI 2020

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER

07. JULI 2020

Datum / Date: