
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW® Apixaban CONT Low 5x1 ml

REF 5090271
LOT 3902C00.01
 2022-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		70.6 – 131.2 ng/mL	100.9	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	70.6 – 131.2 ng/mL	98.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

08. FEB. 2021
Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

08. FEB. 2021
Datum / Date: