



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRK00.01

Att.05

RK11C00.01

TECHNOZYM vWF:CBA Collagen Typ I ELISA Kit 96T.

REF	5450311
LOT	RK11C00.01
	2022-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND02C0	2022-09-30
1x	0.5 mL	vWF Calibrator 2	2ND02C0	2022-09-30
1x	0.5 mL	vWF Calibrator 3	3ND02C0	2022-10-31
1x	0.5 mL	vWF Calibrator 4	4ND02C0	2022-10-31
1x	0.5 mL	vWF Calibrator 5	5ND02C0	2022-09-30
1x	0.5 mL	vWF High Control	AND02C0	2022-10-31
1x	0.5 mL	vWF Low Control	BND02C0	2022-10-31
1x	12 mL	TMB Substrate	RKN02C0	2022-06-30
1x	12 mL	Stop solution	RKK11C0	2023-01-31
1x	80 mL	Wash buffer concentrate	RNA02C0	2022-07-31
1x	90 mL	Incubation buffer	RNB02C0	2022-06-30
1x	-	vWF:CBA Plate Collagen Typ I	RNM01C0	2022-04-30
1x	0.3 mL	CBA Conjugate	RNH01C0	2022-09-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal5	≥ 0.7	2.4	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND20C0	0.80 - 1.34 IU/mL	1.32 IU/mL	OK
	vWF Low Control LOT BND02C0	0.19 - 0.45 IU/mL	0.35 IU/mL	OK



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRK00.01

Att.05

RK11C00.01

Durchgeföhrt von
Prepared by

Michael KAFKA

04. FEB. 2021

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

04. FEB. 2021

Datum / Date: