

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	Antithrombin 1 x 25 IU	EXPIRATION DATE:	2022-08
PRODUCT NUMBER:	81079639		
LOT NUMBER:	N0806562		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Antithrombin Activity	25	30	IU/vial	28	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0029	PASS
Intercept @ 37°C	0.40	0.60		0.44	PASS
r <sup>2</sup>	0.990	--		1.000	PASS

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 9/17/2020 *9-17-20*  
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 9/17/2020 *9/17/20*  
 PRINT SIGNATURE

TITLE: Senior QC Analyst