




# M30 Apoptosense® ELISA

**REF** 10011

Kit Lot No. <b>LOT</b> 10011-009	Expiry  2021-08
--	--

Component	A <sub>450 nm</sub>	<b>LOT</b>		
M30 Standard A	0 U/L	0,136	10011SA- 009 2021-08	
M30 Standard B	75 U/L	0,288	10011SB- 009 2021-08	
M30 Standard C	150 U/L	0,389	10011SC- 009 2021-08	
M30 Standard D	250 U/L	0,685	10011SD- 009 2021-08	
M30 Standard E	500 U/L	1,462	10011SE- 009 2021-08	
M30 Standard F	750 U/L	2,262	10011SF- 009 2021-08	
M30 Standard G	1000 U/L	2,768	10011SG- 009 2021-08	
M5 Coated Microstrips		M5-	011 2023-11	
M30 HRP Conjugate		10011C-	009 2021-08	
M30 Conjugate Dilution Buffer		10011CD-	009 2021-08	
TMB Substrate		VLVT-	007 2024-07	
Stop Solution		VLVSS-	009 2023-08	
Wash Tablet		VLVWT-	005 2023-01	
Controls	Expected value (U/L)	Obtained value (U/L)	<b>LOT</b>	
M30 Control Low	125 ± 25 U/L	135,3U/L	10011CL- 009	2021-08
M30 Control High	650 ± 100 U/L	635,7U/L	10011CH- 009	2021-08

Test date: 25.11.2020	Sign: 
-----------------------	---