

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2023-10
PRODUCT NUMBER:	82141339		
LOT:	N1007813		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.3	PASS

PREPARED BY: Bhumi Vasoya Bhumi Vasoya DATE: 10/30/2020
 PRINT SIGNATURE

TITLE: QC Analyst

APPROVAL SIGNATURE: Donna Tencza Donna Tencza DATE: 10/30/20
 PRINT SIGNATURE

TITLE: Senior QC Analyst