
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320
LOT 1Z02C00.01
 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z02C00	2021-05-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA RC Low (nM Peak Thrombin) LOT 7D01C00	170 - 283	189.3 OK
	TGA RC High (nM Peak Thrombin) LOT 7F92B00	458 - 764	679.1 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. DEZ. 2020

Datum / Date:

**Freigabe von
Approved by**

Christoph Enne
Christoph ENNE

09. DEZ. 2020

Datum / Date: