



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0100.01

TECHNOVIEW Orgaran® Calibrator Set 5x1 mL

REF 5090110

LOT 0101C00.01

2022-04-30

Inhalt Content	Reagenz Reagent	LOT	
1x1 mL	TECHNOVIEW Orgaran® CAL1	5L01C00	2022-04-30
1x1 mL	TECHNOVIEW Orgaran® CAL2	5M01C00	2022-04-30
1x1 mL	TECHNOVIEW Orgaran® CAL3	5N01C00	2022-04-30
1x1 mL	TECHNOVIEW Orgaran® CAL4	5O01C00	2022-04-30
1x1 mL	TECHNOVIEW Orgaran® CAL5	5P01C00	2022-04-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Orgaran® Cont. Low LOT 3Q81B01	0.43 – 0.79 IU/mL	0.63 OK
	TECHNOVIEW Orgaran® Cont. High LOT 3W01C00	0.77 – 1.28 IU/mL	1.06 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Stefan Steiner

03. FEB. 2021

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

04. FEB. 2021

Datum / Date: