



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7PT00AZ

TECHNOZYM[®] ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463-RUO

LOT 7PT11C0.02

 2022-09-30

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|---------------------------|---------|---|
| 1 x 0.5mL | Adamts13 INH High Control | RPY01C0 | 2022-09-30 |
| 1 x 0.5mL | Adamts13 INH Low Control | RPZ01C0 | 2022-09-30 |

| Analytische Prozedur Analytical Procedure | Anforderung Requirement | Ergebnis Result | | |
|--|----------------------------|--------------------|-------|----|
| Verwendbarkeit Usability | High Control (RPY01C0) | 54.04 – 90.06 U/mL | 79.25 | OK |
| | Low Control (RPZ01C0) | 14.50 – 33.82 U/mL | 27.24 | OK |

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

10. MRZ. 2021

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

11. MRZ. 2021

Datum / Date: