

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2023-01
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N0119671	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
Analysis with COATEST APC R:						
APT Time	<u>CaCl₂</u>	27	44	seconds	44	Pass
	<u>APC/CaCl₂</u>	--	101	seconds	92	Pass
APC Ratio		--	2.3	--	2.1	Pass

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	27	40	seconds	34	Pass
APC Ratio		1.4	1.9	--	1.8	Pass

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: D. Tencza *D. Tencza*
 PRINT SIGNATURE
 TITLE: Senior QC Analyst

DATE: 1/25/2021 *1/25/21*

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*
 PRINT SIGNATURE
 TITLE: Technical Supervisor

DATE: 1/25/2021 *1/25/21*