

# CERTIFICATE OF ANALYSIS



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2023-02
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0210347	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.34	PASS
Intercept	---	0.12	--	0.06	PASS
r	0.990	---	--	0.995	PASS
LOW RANGE					
Slope	2.8	6.9	--	3.4	PASS
Intercept	---	0.21	--	0.09	PASS
r	0.990	---	--	0.999	PASS

**KIT COMPONENTS**

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0210345	3 mL
COAMATIC FVIII S-2765 + I-2581	N0210346	6 mL
COAMATIC FVIII BUFFER	N0210342	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N.M. Patel*  
 PRINT SIGNATURE

DATE: 2-19-21  
 2/19/2021

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*  
 PRINT SIGNATURE

DATE: 2/19/21  
 2/19/2021

TITLE: Technical Supervisor