

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3V00.01
---	---	-----------

**TECHNOVIEW Arixtra CONT H 6x1 mL**

**REF** 5090014-RUO

**LOT** 3V02C00.02

 2022-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa Arixtra	1.06 – 1.76 µg/mL	1.41	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa Arixtra	1.06 – 1.76 µg/mL	1.37	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Christoph ENNE

29. APR. 2021

Datum / Date:

**Freigabe von**  
**Approved by**

  
Michael KAFKA

29. APR. 2021

Datum / Date: