
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0H00.01
---	---	-----------

TECHNOVIEW LMW CAL Set 5x1 mL

REF 5090040-RUO

LOT 0H11C00.02

 2022-11-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW LMW Cal 1	5A11C00	2023-02-28
1x1mL	TECHNOVIEW LMW Cal 2	5B01C00	2022-11-30
1x1mL	TECHNOVIEW LMW Cal 3	5C01C00	2022-11-30
1x1mL	TECHNOVIEW LMW Cal 4	5D01C00	2022-11-30
1x1mL	TECHNOVIEW LMW Cal 5	5E01C00	2022-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW LMW Cont L LOT 3F91B00	0.17 – 0.31 IU/mL	0.30	OK
	TECHNOVIEW LMW Cont H LOT 3H91B00	0.83 – 1.39 IU/mL	0.99	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas HOLOVICZ

29. APR. 2021

Datum / Date:

Freigabe von
Approved by


GRESAK PODAKAR

29. APR. 2021

Datum / Date: