

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3Q00.01
---	---	-----------

TECHNOVIEW® Organan CONT Low 5x1 mL

REF 5090112-RUO

LOT 3Q01C00.02

 2022-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.38 – 0.70 IU/mL	0.54	OK
Verwendbarkeit Usability Ceveron® alpha	Technochrom® anti Xa Organan	0.38 – 0.70 IU/mL	0.56	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

29. APR. 2021

Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

29. APR. 2021

Datum / Date: