

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3300.01
---	---	-----------

## TECHNOVIEW® Rivaroxaban Control High 5x1 mL

**REF** 5090174-RUO

**LOT** 3301C00.02

 2022-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		234.7 – 391.1 ng/mL	312.9	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW® anti-Xa	234.7 – 391.1 ng/mL	298.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

**Lucas HOLOVICZ**



29 APR 2021  
Datum / Date:

**Freigabe von**  
**Approved by**

**Suzana GRESAK PODAKAR**



29. APR. 2021  
Datum / Date: