



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3900.01

TECHNOVIEW® Apixaban CONT Low 5x1 ml

REF 5090271-RUO

LOT 3902C00.02

 2022-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		70.6 – 131.2 ng/mL	100.9	OK
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	70.6 – 131.2 ng/mL	98.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Fikret NASUFI 

29. APR. 2021  
Datum / Date:

Freigabe von  
Approved by

 Christoph ENNF

29. APR. 2021  
Datum / Date: