
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0W00.01
---	---	-----------

**Technochrom AT III Kit ~ 100 T.**

REF      5340224-RUO

LOT      0W04C00.03

      2023-07-31

Inhalt Content	Reagenz Reagent	LOT	
2x25 mL	NaCl Solution 0,9 %	9591B00	2024-02-29
1x	ATIII Substrate Th-1 10 µmol	8B03C00	2023-07-31
1x	Technochrom ATIII Reagent A2 20 µmol	7S01C00	2023-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha Chromogenic Substrate F IIa	CCN LOT 1P02C00	75.8 – 126.3 %	102.9	OK
	CCA LOT 3P91B00	33.8 – 56.3 %	46.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
Suzana GRESAK PODAKAR

29. APR. 2021

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA


29. APR. 2021

Datum / Date: