



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQRR00AZ

TECHNOZYM[®] ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601-RUO

LOT RR12C00.01

 2022-06-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS02C0	2022-12-31
1 x 0.3mL	Adamts13 Ag Conjugate	RQV02C0	2022-12-31
1x 80mL	Wash buffer Concentrate	RNA04C0	2022-10-31
1 x 90mL	Incubation buffer	RNB03C0	2022-10-31
1 x 12mL	TMB Substrate	RKN02C0	2022-06-30
1 x 12mL	Stop solution	RKK11C0	2023-01-31
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag High Control	RRY01C0	2022-10-31
1 x 0.5mL	Adamts13 Ag Low Control	RRZ01C0	2022-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.87	OK
Verwendbarkeit Usability	High Control (RRY01C0)	0.47 – 0.79 IU/mL	0.55	OK
	Low Control (RRZ01C0)	0.12 – 0.28 IU/mL	0.18	OK

Durchführung von
Prepared by

Fikret NASUFI 

09. FEB. 2021

Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

09. FEB. 2021

Datum / Date: