


FB 1.2.3.93	Certificate of Analysis	
-------------	-------------------------	---

Certificate of Analysis

Product Name: Senova Von Willebrand Factor Propeptide ELISA Test Kit (ELISA VWF:PP)

LOT/SID: 2.7.2.17-6

Intended Use: quantitative measurement of von-Willebrand-factor-propeptide in human blood plasma

Analyte: von-Willebrand-factor-propeptide

Probe matrices: human blood plasma

Shelf life and storage: 12 months from date of manufacture at 2-8°C

Expiry date: 2022-05

Characteristics:

Detection threshold	2.5 mIU/ml
Lowest detectable concentration	5 mIU/ml
Measuring range	2.5 – 120 mIU/ml

Approved by Quality Assurance: 2021-06-03 A. Jonek

Electronic signature

For research use only