



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQB400AZ

Ceveron® TGA RC Low Kit 120T

REF 5006013-RUO

LOT B411C00.02

🕒 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	🕒
2x1 mL	Ceveron® TGA RC Low	7R02C	2022-07-31
2x1.5 mL	Ceveron® TGA BUF	9Q91B	2021-11-30
2x3 mL	Ceveron® TGA SUB	8E01C00	2023-01-31
2x1 mL	Ceveron® TGA Control High	1Y03C00	2022-02-28
2x1 mL	Ceveron® TGA Control Low	2703C00	2022-02-28
1x25 mL	Ca-Chloridlösung 25 mmol/L	9C03CB0	2025-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	Ceveron® TGA Control-High (nM Peak Thrombin) LOT 1Y03C00	200 – 333	271.9	OK
	Ceveron® TGA Control-Low (nM Peak Thrombin) LOT 2703C00	69 – 153	128.3	OK
Etiketten und Barcode Überprüfung Label and bar code verification			OK	

Durchführung von
Prepared by

Fikret NASUFI

27. MAI 2021
Datum / Date:

Freigabe von
Approved by

Michael KAFKA

27. MAI 2021
Datum / Date: