

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	0201C00.04
---	---	------------

Technoview® Argatroban CAL Set 5x1mL

REF	5090140-RUO
LOT	0201C00.04
	2022-06-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT		
			0201C00		
1x	1mL	Argatroban CAL 1	5Q51B01		2022-06-30
1x	1mL	Argatroban CAL 2	5R51B01		2022-06-30
1x	1mL	Argatroban CAL 3	5S51B01		2022-06-30
1x	1mL	Argatroban CAL 4	5U51B01		2022-06-30
1x	1mL	Argatroban CAL 5	5V51B01		2022-06-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
	Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	Technoview® Argatroban CONT L LOT 3151B01	0.48 - 0.90 µg/mL	0.70 µg/mL
	Technoview® Argatroban CONT H LOT 3251B01	0.90 - 1.50 µg/mL	1.10 µg/mL	OK

Durchgeführt von
Prepared by


Lucas HOLOVICZ

15 JUNI 2021

Datum / Date:

Freigabe von
Approved by


Suzana GRESAK PODAKAR

16 JUNI 2021

Datum / Date: